



CE DECLARATION OF CONFORMITY

We,

Nedis B.V.
De Tweeling 28
5215MC 's-Hertogenbosch
The Netherlands
Tel.: 0031 73 599 1055
Fax.: 0031 73 599 9699
Email: info@nedis.com
Internet: www.nedis.com

Declare under our responsibility that the product;


Brandname: König Electronic
Model: HC-BLDPRESS10
Description: Electronic blood pressure meter

Is in conformity with the following standards;
EN 1060 -1: 1995 +A1: 2002, EN 1060 -3: 1997 +A1: 2005
IEC 60601 -1: 1988 +A1: 1991 +A2: 1995
EN 1060 -4: 2004, EN 980: 2003, EN 1041: 1998
EN 60601 -1 -2: 2007, EN ISO 14971: 2007
EN ISO 10993 -1: 2003, EN ISO 10993 -5: 1999
EN ISO 10993 -10: 2002, BN EN 60601 -1 -6: 2004

Following the provisions of the MDD 93/42/EEC Directive.

Produced conform: EN ISO 9001: 2008

's-Hertogenbosch, 29-6-2009


NEDIS B.V.
Postbus 70509, 5201 GL 's-Hertogenbosch NL
De Tweeling 28, 5215 MC 's-Hertogenbosch NL
Telefoon: 073 - 599 96 41

Mrs. J. Gilad
Purchase Director

CE