



## CE DECLARATION OF CONFORMITY

We,

Nedis B.V.  
De Tweeling 28  
5215MC 's-Hertogenbosch  
The Netherlands  
Tel.: 0031 73 599 1055  
Fax.: 0031 73 599 9699  
Email: [info@nedis.com](mailto:info@nedis.com)  
Internet: [www.nedis.com](http://www.nedis.com)

Declare under our responsibility that the product;

Brandname: König  
Model : HC-BLDPRESS20  
Description: Automatic Blood Pressure Monitor

Is in conformity with the following standards;

EN980: 2008, EN1041: 2008, EN ISO 14971: 2007, EN62366: 2008,  
EN62304: 2006, EN ISO 10993-1: 2003, EN ISO 10993-5: 1999,  
EN ISO 10993-10: 2002/A1: 2006, EN60601-1: 1990+A1+A2,  
EN60601-1-2: 2001+A1, EN60601-1-4: 1996+A1, EN60601-1-6: 2004,  
EN1060-1: 1995/A1: 2002, EN1060-3: 1997/A1: 2005, EN1060-4: 2004  
EN ISO 9001: 2008

Following the provisions of the MDD 93/42/EEC Directive.

Produced conform: EN ISO 9001: 2008

's-Hertogenbosch, 08-03-2010

NEDIS B.V.

Postbus 70509, 5201 CL 's-Hertogenbosch NL  
De Tweeling 28, 5215 MC 's-Hertogenbosch NL  
Telefoon: 073 599 9641

Mrs. J. Gilad  
Purchase Director

CE